Journal of Applied Health Sciences and Medicine, 1 (2) (2021) 28-37

A detection of C-reactive protein (CRP) using laser

Madhanlal Udhayarasu[1]

Received: 18 September 2021 Accepted: 15 November 2021

Abstract

A laser-based approach for detecting C-reactive protein (CRP) is proposed in this study. Laser beams are devices that generate powerful monochromatic, coherent, and heavily collimated beams of light. When in comparison to other light references, the range of frequencies (colour) of photons from a laser is exceptionally solid (monotonous). C-reactive protein or CRP is a pentameric, protein with an annular shape present in plasma from the bloodstream that monitors overall levels of inflammation in the body. It got its name since it was initially discovered in the serum of patients suffering from acute inflammation. Infections and numerous long-term disorders create high levels of CRP. Burns, trauma, certain cancers are the main causes of elevated c reactive protein. Its goal is to characterise and optimise the technique of coating CdSe quantum dots suspended in ZnS to considerably increase their quantum yield. To effectively study the effects of a ZnS covering on the quantum dots, a current process for synthesising, covering, and evaluating CdSe quantum dots had to be developed.

Keywords: LASER (650 nm), Photo transistor, Dots of quantum information, lights, biological imaging, fluorescence, absorbance, spectrophotometer, spectroscopy.

1. Introduction

C-reactive proteins (CRP) is a pentameric, glycoprotein with an annular shape present in plasma from the blood that rises in reaction to inflammatory. CRP was named after a material seen in the blood serum suffering from severe irritation that responded with the Pneumococcus C-polysaccharide. It was a blood test used to determine how much C-reactive protein is present in our blood. C-reactive protein (CRP) is a blood test that determines how

[1]Department of Medical Electronics, Saveetha Engineering College, Chennai - 602117, India
*Email: madhanlal@saveetha.ac.in

much inflammation is present in our bodies. Infections and numerous long-term disorders create high levels of CRP. However, a CRP test can't tell you where the inflammation is or what's producing it. To determine the source and location of the inflammation, more tests are required. In reaction to substances released by macrophages and fat cells, the liver produces CRP (adipocytes). Pentraxin is a protein that belongs to the pentraxin family. This has nothing to do with C-peptide (insulin) or protein C (blood clotting) [1]. CRP is an intense protein with a molecule that is homopentameric and phosphocholine-specific Ca+2 binding specificity or PCh. This is a member of the lectin folded group and is a nonglycosylated circulating member of the pentraxin family. Pentraxins are a pattern-recognition protein family made composed of five identical subunits that have evolved over time. The pentraxins are split into two classes depending on the subunit's structural system: There are two types of pentraxins: Pentraxins are divided into two types: brief & medium pentraxins. The two short pentraxins are CRP and SAP. Pentraxin 3 is the long pentraxin group's prototypical protein (PTX3).

CRP is a chemical found in humans made up of five nonglycosylated polypeptide subunits that are noncovalently coupled, grouped in cyclic pentameric symmetry, and structured in a disc-like arrangement around a central pore. The mass of each component is 23,027 Da (206 amino acid residues), while the total mass of the human CRP molecule is 115,135 Da. Human CRP receptors that have been generated are differently glycosylated with various quantities of sialic acid derivatives with particular linkages and fluctuate in melatonin content in different clinical situations, resulting in a molecular weight fluctuation of around 2000–4000 Da. [2]. The liver has long been known to produce CRP as a reaction to inflamed mediators. CRP can be produced by nonhepatic tissues, according to several recent investigations. Under some conditions, both pulmonary epithelial cells and renal epithelial cells can create CRP, according to two studies. Neuronal cells also appear to have the ability of producing involved in the reaction in the acute stage that are involved in the pathophysiology of neurodegenerative diseases like Alzheimer's. Several additional CRP-producing sources pointed to another more systemic CRP creation in our bodies. These additional sources, however, only offered a shaky relationship to atherosclerosis. In atherosclerotic plaques, CRP has been found to colocalize with the terminal component group. CRP generated locally may have a in the pathogenesis of atherosclerosis [2,3]. Furthermore, we discovered that human adipocytes can create CRP in response to cytokines that cause inflammation and a particular adipokine called resistin. Adipocytes' development of CRP may help to describe why CRP levels are high in people who have insulin resistance (Fig1).



Figure 1: The liver, adipocytes, and smooth muscle cells of the vascular system all create CRP

Contents of C reactive protein in the standard range:

CRP: 0-10 mg/dL is the C-reactive protein standard region, CRP with high sensitivity (hs-CRP): 3 mg/L. CRP levels in the blood rise as a result of an inflamed disorder in the body. Inflammation is not an issue in and of itself, but it can signal a variety of all of the other health issues, Infections, arthritis, kidney failure, and pancreatitis are just a few examples. Patients with CRP levels that are too high are a greater chance of developing coronary heart disease, which can lead to a heart attack. The American Heart Association and The CDC (Centers for Disease Control and Prevention) is a federal agency that works to prevent disease issued clinical recommendations on CRP testing in 2003. They assigned CRP values to determine levels of cardiovascular risk. [4] Reduced risk would be less than 1.0 mg/L, intermediate risk is 1.0-3.0 mg/L, and high risk is greater than 3.0 mg/L.

The JUPITER study has a significant consequence on the utility of hs-CRP as a risk marker. The study finishes with an evidence-based conclusion about the present role of hs-CRP in anticipating CV risk. The usage of C-reactive protein with a great specificity as a measure of cardiovascular (CV) risk is fraught with controversy. Even though there is a lot of evidence linking hs-CRP to cardiovascular illness, its role in clinical practise is still unknown. A medical announcement from the American Heart Association (AHA) was recently published on criteria for evaluating novel cardiovascular risk markers [5]. C-reactive an antigen and LDL cholesterol were assessed at the start of the study in 27,939, seemingly healthy Women in the america, who were then followed of an average of eight years for myocardial injury, ischemic stroke, coronary angioplasty, or mortality from cardiopulmonary disease. However, there was only a minor correlation (r=0.08) between C-reactive protein and LDL cholesterol,

baseline levels of both There was a significant linear link between them with the incidence of risk of heart disease [6]. These obtained results are found to be acceptable with the clinical results.

2. Materials and Method

2.1. Study Population and Data Collection

To enable exclusive detection of CRP, a set of test QDs (CdSe and CdSe/ZnS) were synthesised utilising high-purity sodium chloride (NaCl2), distilled water, potassium (k), and charcoal for QD coating.

The CdSe quantum dots were synthesised, then characterised before being covered with ZnS and characterised subsequently. Implementing this approach allowed for direct comparisons between coated and untreated quantum dots, allowing the effects of the ZnS surface coating to be characterised.

Sodium chloride and potassium chloride were used to make CdSe QDs. Core-shell quantum dots made of cadmium selenide and zinc sulphide, cadmium selenide (CdSe) quantum dots were produced. This will show the steps involved in creating CdSe quantum dots. The resulting solution was placed in a cuvette, and then the laser light was used to pass between the regular solution and sodium carbonate.

Lasers are devices that generate powerful monochromatic, coherent, and extremely collimated beams of light. As opposed to other light sources, laser light has a shorter wavelength (colour) is exceedingly monochromatic, and the total amount of energy in the laser beam have a set association between phases with others. The divergence of laser light is often quite minimal. It has the ability to long distance transportation or to be focussed on a small area with a brightness greater than that of the sun. Lasers are employed in a wide range of applications due to their properties. This light is viewed using a LASER with a wavelength of 650 nm and a power of 5 mW.

2.2 LASER (Stimulated emission of radiation amplifies light)

Stimulated Emission of Radiation Amplification of Light. Lasers are devices that generate powerful monochromatic, coherent, and heavily collimated beams of light.

Figure 2: A 650 nm at LASER

A laser's operation is dependent on three distinct characteristics:
a) Inside of an amplifying medium, stimulated emission
b) Inversion of the populace due to electronic devices and
c) A modulator for light

2.3. TURBIDIMETER

On horse blood samples, a turbidimetric approach determined by the interplay of soybean oil and phosphocholine was used to measure serum CRP concentration in order to assess its a possible diagnosis utility in medical care for animals. A mixture of horse serum and 20% intralipid in a Tris-calcium buffer of 0.1 M (pH 7.5). The CRP-phosphocholine compounds were bichromatically turbidimetrically measured. (660 nm/700 nm) measured using a analyser for business after 30 minutes of incubation at 37 °C. Furthermore, the therapeutic utility including both tests was evaluated by comparing CRP to other inflammation markers such like white blood cell and neutrophil counts. A standard human CRP calibrator has been used to standardise the test. CRP values were taken from blood serum (296 patients and 34 controls). The standard concentrations were observed to be less than ten milligrammes per litre. Among 1 and 400 mg/l, the technique was demonstrated toward being linear. CRP levels and relative neutrophil counts were found to have a modest association. The area under the curve for CRP was0.928, and that was the best (P0.001) to the neutrophil count was 0.804 at the time and the leukocyte count(0.664) when it comes to spotting the existence of inflammatory response, according to receiver-operating characteristics analysis. As such an Horses are put through an acute phase exam, this CRP assay produced consistent results, demonstrating it is not dependant on the breed capacity to recognise CRP in a variety of mammals, especially horses.



Figure 3: Instruments for turbidimeters

2.4. To make the QD-antibody complex, follow these steps.

The QD-mAb combo was synthesised according to Biosystems' guidelines. To make the QD-antibody complex, the QD-streptavidin conjugate and the monoclonal antibody were combined and divided at 25°C for 2 hours with moderate stirring. The QD-mAb combination was then diluted and filtered before the transmittance and fluorescence spectra for the discharge were collected. Excess antibodies were found in samples with 280 nm sensitivity and no luminescence at the QDs' release optimum. That pairing efficiency was calculated using a Bradford assay, which measured the quantity of detached antibody. By deducting the total of antibodies that is detached from the overall the overall amount of antigens employed in the testing, the conjugated antibody content was calculated. This QD-antibody conjugate's produced concentrations were calculated appropriately.

2.5 Cdse/Zns coated Quantum dots

Because of their unusual optical features, quantum dots (QD) nanoparticles have been widely exploited in biomedical and electronics areas. As a result, it has a huge risk of human exposure and ecosystem discharge. A variety of protective coating or chemical bonding are applied to QDs to boost their bioactivity and water solubility, hence increasing their biocompatibility.

The ZnS layer all around CdSe quantum dots protects and stabilises the quantum dots' superior optical characteristics against harmful environmental influences. The produced core

nanocrystals have the emission spectra characteristics are identical as shell CdSe quantum dots, although with higher quantities of brighter fluorescence, higher planning process typically and durability, and better interface chemistry control.



Fig 4: Cdse/Zns visible Quantum dots -water soluble -Amine coating

A laser-based approach for detecting C-reactive protein (CRP) has been developed. Lasers are devices that generate powerful monochromatic, coherent, and extremely collimated light beams. When in comparison with another light resources,of light is exceptionally pure. CRP is a pentameric, circumferential protein present in blood plasma that monitors overall levels of chronic inflammation. It got its name since it was initially discovered in the serum of patients suffering from acute inflammation. Infections and numerous long-term disorders create high levels of CRP. The most common causes of increased C reactive protein are burns, trauma, and some malignancies. Its goal is to characterise and optimise the technique of coating CdSe quantum dots contained in ZnS to considerably increase their quantum efficiency. Quantum dots are extremely tiny semiconductor crystals. They varied in diameter from a few hundred to a few thousand atoms. A quantum dot has a size of around two to ten nanometers. A process flow for synthesising, coating, and characterising CdSe quantum dots required to be created in order to reliably examine the effects of a ZnS coating on the quantum dots. The CdSe quantum dots were synthesised, then characterised before being coated with ZnS and characterised again.

This approach allows us to compare coated and uncoated quantum dots, allowing us to characterise the impacts of the ZnS coating process.



Figure 5: CRP detection setup block diagram

As a light source, For the detection of C reactive protein, LASERs with a wavelength range of 650nm are used (CRP), it is passed through a tube which contains CRP. Suction devices are used to take the accurate reading to clear airways of materials that would impede breathing or cause infections. The CRP serum is passed through a detector.

3. Results

This study is able to optimize the value of liquid (distilled water) with Nacl solution or without Nacl solution. It is examined with the help of a LASER and numerous characterisation components such as photo transistors, resistors, and integrated circuits (integrated circuit). The inference obtained with that tests are presented below:

CONDITION	VOLTAGE
Normal environment (dark)	**1.46**
Normal environment (bright)	**1.50**
Normal environment (closed)	**1.02**
Fully closed	**0.96**

Table 1: without sample

CONDITION	VOLTAGE
Normal sample (distilled water)	**1.20**
With opaque object	**0.80**
With Nacl	**1.17**

Table 2: Different sample sizes were used in comparative research.

4. Summary and Future work

A detector is used to show it for miligram quantities of various materials such as Nacl and regular freshwater. The procedure used to define the sensor for the detection of samples. It is examined utilising a Lasers, a photo diode, a resistor, and an integrated circuit.

Measurement of CRP by Synthesisation and coating in the ZnS reaction must be done using quantum dots in this study. To acquire a better understanding of how the process might be improved, the quantum dots created by this and related approaches will be characterised. This quantum dots produced is found to have possible application in other field.

References

[1] T T Abd, D J Eapen, A Bajpai, A Goyal, A Dollar, L Sperling. The role of C-reactive protein as a risk predictor of coronary atherosclerosis: implications from the JUPITER trial. Curr Atheroscler Rep, 13:154-161, 2011.
[2] S Black, I Kushner, D Samols. C-reactive protein. J Biol Chem, 279:48487-48490, 2004.
[3] Coe-Sullivan, Seth. Hybrid organic/quantum dot thin film structures and devices. PhD thesis, Massachusetts Institute of Technology, 2005.
[4] H Harma, J Toivonen. Time-resolved fluorescence immunoassay for C-reactive protein using colloidal semiconducting nanoparticles. Sensors, 11:11335-11342, 2011.
[5] I J W Park, S. Kurosawa, D S. Han, H. Aizawa, M. Yoshimoto, C. Nakamura, J. Miyake, S. M. Chang. “Environmental immunosensor detection for 2,4-dinitrophenol as a model compound of dioxin”, I’rweedingv of (he 200 I IEFX international Frequency Control Symposium, pp.489-491, 2008.
[6] M D S Islam, S H Kang. Chemiluminescence detection of label-free C-reactive protein based on catalytic activity of gold nanoparticles. Talanta, 84:752–758, 2011.
[7] V K Prashant. Quantum Dot Solar Cells. Semiconductor Nanocrystals as Light Harvesters. The Journal of Physical Chemistry C, 112:18737-18753, 2008.
[8] J M Catherine, J L Coffer. Quantum Dots: A Primer. Appl Spectrosc, 56, 16-27, 2002.
[9] B M Christopher. Synthesis and Characterization of II-VI Quantum Dots and Their Assembly into 3D Quantum Dot Superlattices. PhD thesis, Massachusetts Institute of Technology, 1995.

[10] T A Pearson, G A Mensah, Y Hong, S C Smith Jr. CDC/AHA workshop on markers of inflammation and cardiovascular disease: application to clinical and public health practice: overview. Circulation, 110:543-544, 2004.
[11] P M Ridker, N Rifai, L Rose, J E Buring, N R Cook. Comparison of C-reactive protein and low-density lipoprotein cholesterol levels in the prediction of first cardiovascular events. New England J Med, 347:1557-1565, 2002.
[12] R Peter, M Protière, L Li. Core/Shell Semiconductor Nanocrystals. Small, 5:154-168, 2009.
[13] R G Ute, M Grabolle, S C Jaricot, R Nitschke, T Nann. Quantum dots versus organic dyes as fluorescent labels. Nature Methods, 5:763-775, 2008.
[14] Rogach, L Andrey. Semiconductor nanocrystal quantum dots: synthesis, assembly, spectroscopy and applications. Wien, Springer, 2008.
[15] S. Reghuram, A. Arivarasan, R. Kalpana and R. Jayavel, Department of Biomedical Engineering, Rajalakshmi Engineering College, Chennai 602105, India; Centre for Nanoscience and Technology, Anna University, Chennai 600025, CdSe and CdSe/ZnS quantum dots for the detection of C-reactive protein India [(Received 27 June 2013; final version received 1 March 2014)].
[16] S S Jonathan, P Snee, S Coe-Sullivan, J P. Zimmer, J E. Halpert, P Anikeeva, L Kim, V Bulovic, M G. Bawendi. Color-Saturated Green-Emitting QD-LEDs. Angewandte Chemie International Edition 45, 5796-5799, 2006.
[17] P L Tugirimana, A L Holderbeke, J A K Dela. A new turbidimetric method for assaying serum C-reactive protein based on phosphocholine interaction, 47:1417-1422, 2009.
[18] J V Veetil, M Mehta, W Andrew, K Ye, Quantum Dot based lateral dot strips assays Technical conference, IEEE, 2007.
[19] Y. N. Yang, H. I. Lin, J. H. Wang, S. C. Shiesh and G. B. Lee. An integrated microfluidic system for C-reactive protein measurement. Biosens. Bioelectron, 24:3091-3096, 2009.
[20] Y W William, L Qu, W Guo, X Peng. Experimental Determination of the Extinction Coefficient of CdTe, CdSe, and CdS Nanocrystals. Chemistry of Materials, 15:2854-2860, 2003.